



lot # 42

1935 Plaza Real
Oceanside, CA 92056-6024

National Test Equipment, Inc.
www.nationaltestequipment.com

Ph: (760) 639.1
Fx: (760) 639.1

Certificate of Calibration

ANSI/NCSL Z540-1-1994

Certificate Number: 527965-1



Model: Tektronix TDS3054
Description: 500MHz 4CH 5GS/s Color Digital Phosphor Oscilloscope
Serial No: [Redacted]
Asset No: N/A

Date of Calibration: Nov 29, 2021
Date Calibration Due: Nov 29, 2021 (0 Months)
Temperature: 70.68 F°
Humidity: 33.25 %
Procedure: METCal
Technician: # 123

Customer:



As Received: **In-Specification** - Measured values were within specifications at the points tested.

Actions Taken: None

As Returned: **In-Specification** - Measured values were within specifications at the points tested.

Calibration Data: (See Attached)

Standards Used:

Fluke 9500B 103856 Due 4/9/22
Fluke 9530 109901 Due 4/23/22

Fluke 9530 118127 Due 5/18/22
Fluke 9530 109899 Due 6/11/22

Fluke 9530 109900 Due 4/24/22

The equipment listed above meets or exceeds published specifications and has been calibrated in compliance with ANSI/NCSL Z540-1-1994 measurement standards whose accuracies are traceable to the National Institute of Standards and Technology (NIST) or any other recognized National Metrology Institute (NMI). National Test Equipment's measurement standards calibration system is compliant ANSI/NCSL Z540-1-1994 and MIL-STD 45662A.

Traceability is to the International System of Units (SI), consensus standards or ratio type measurements through national standards realized and maintained by the NIST or any other NMI.

This calibration applies only to the unit listed on this certificate. This certification shall not be reproduced except in full, without the written approval of National Test Equipment, Inc.

Laboratory Manager:

Steve Clifton

Document Print Date Dec 02, 2021

Form NTECCAL-0002